

CUSTOMER		SHIP TO	
Company		Name	
Account #	Date	Address	
<input type="checkbox"/> Quote	<input type="checkbox"/> PO #	City	Province
Buyer		Post Code	Phone
Mark For		Email	
<input type="checkbox"/> Quote has Additional Sunrise Products with different Quote / Order #		Quote / Order #	

**Technical Data Overview**

Maximum User Weight	242 lbs (110 kg)	Max Speed	9.3 mph (15 kph)
Front Frame Width	10.2 - 18.9 in (260 - 480 mm)	Motor Wheel Size	6.5 in Solid Tire
Total Weight with Subframe	From 19.8 lbs (9 kg)	Brake Type	Dual Brake System, Mechanical and Motor
Battery Type	Li-Ion, 36 V, 4 Ah	Operating Voltage	36V
Range	9.3 mi (15 km)	Motor Power	6.5 in 250 watt (nominal)
Charger	36V, 2A	Climbing Power	Up to 11% (6°)

**Empulse F35 Configuration**

<input checked="" type="checkbox"/> OPTION	Description	RETAIL
<input type="checkbox"/> 147MW95	Empulse F35	F35 Stick, Subframe mount, Frame clamps kit, and Charger
Max Speed -		\$ 4,825.00

**Select only one option in this section.**

The max speed of the F35 will be limited to the selected speed.

<input type="checkbox"/> 147S104	9.3 mph (15 kph)	NC
<input type="checkbox"/> 147S105	3.7 mph (6 kph)	NC

**Options and Accessories**

<input type="checkbox"/> 147BJ37	Additional Charger	\$ 250.00
----------------------------------	--------------------	-----------

**Wheelchair Docking**

<input type="checkbox"/> 1479A27	Docking for Fixed-Front Wheelchairs	NC
<input type="checkbox"/> 1479A28	Docking for Swing-Away Wheelchairs	NC

**Docking Clamps**
**Select only one option in this section**

Check installation space for the docking clamps before ordering. There needs to be at least 1.5 in (35 mm) for the clamp to attach on the wheelchair's frame front.

**Sunrise Medical Chairs**

<input type="checkbox"/> 1479V58	Quickie Nitrum	<input type="checkbox"/> 1479V59	Quickie QRi	<input type="checkbox"/> 1479V79	Quickie 5R	<input type="checkbox"/> 1479V68	Quickie 2
<input type="checkbox"/> 1479V69	Quickie QS5 X	<input type="checkbox"/> 1479V60	Quickie QX / QXi	NOTE: QXi is only compatible with the standard frame configuration			


**Standard Clamp Kits**

<input type="checkbox"/> 147AC3	Docking Clamps for Oval Tubes	NC
<input type="checkbox"/> 147AC5	Docking Clamps for Round Tube Frames with Diameter of 0.76, 0.9, 0.98, 1.1, and 1.2 in (19.5, 23, 25, 28.6, and 30 mm)	NC
<input type="checkbox"/> 147AC6	Docking Clamps for Round Tube Frames with Diameter of 1.25 in (32 mm)	NC

**Front Frame Width**
**Select only one option in this section.** Please order the smaller version when a distance is measured between mentioned values.

Measure from center frame tube to center frame tube at the mounting position. Use only the measured dimension and not the seat width to check compatibility.

<input type="checkbox"/> 147W61	11 in (279 mm)	<input type="checkbox"/> 147W67	17 in (432 mm)
<input type="checkbox"/> 147W62	12 in (305 mm)	<input type="checkbox"/> 147W68	18 in (457 mm)
<input type="checkbox"/> 147W63	13 in (330 mm)	<input type="checkbox"/> 147W69	19 in (483 mm)
<input type="checkbox"/> 147W64	14 in (356 mm)		
<input type="checkbox"/> 147W65	15 in (381 mm)		
<input type="checkbox"/> 147W66	16 in (406 mm)		



To find the distance between the centerlines of the front frame tubes, **measure the wheelchair directly** if possible. If the chair is not available, you can calculate this distance for either rigid or folding frames by **subtracting the inset** from each side of the seat width, then **subtracting the frame tube diameter**.

**Example**

Chair Configuration: Seat width: 17.5" | Inset: 1" (per side) | Frame tube width: 1.0"

Calculated Front Frame Width is 14.5".

Order From Selection is 14", the smaller size selected is calculation is in between sizes.